



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer The University of Alaska, Fairbanks  
Job Number 73739  
Date 3/27/2013  
Technician CJ

Serial Number 051817

Low Pressure (PSI) 45 PSI  
Time (Minutes) 15 Minutes

High Pressure (PSI) 10500 PSI  
Time (Minutes) 30 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.

