



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer The University of Alaska, Fairbanks
Job Number 73739
Date 3/27/2013
Technician CJ

Serial Number 052222

Low Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 10500 PSI
Time (Minutes) 30 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.

