

# WindObserver II™ Product Test Report



Product Tested: WindObserver II  
Part Number: 1390-75-B-313  
Serial Number: 1351005 - WC45  
Test Date: 20/12/2013  
Location: Gill Instruments Ltd

GILL ensures that quality is inherent in all aspects of the activities and ensures that compliance with BS EN ISO9001: 2008 is maintained.

This report certifies that the above instrument has been tested in accordance with Gill internal procedures

## Results

Test	Limits	Passed
Still Air Test (Zero Wind Speed)	< 0.02m/s	Pass
Wind Tunnel Test (12 m/s nominal)	Pass/Fail	Pass

Generic calibration is traceable to the University of Southampton wind tunnel and instrumentation is maintained in accordance with UKAS.

All tests have been successfully completed

On behalf of Gill Instruments Ltd

Tony Aspland-Jupe  
Quality Control

2002-0395 Issu



Gill Instruments Ltd  
Saltmarsh Park  
67 Gosport Street  
Lymington  
Hampshire  
SO41 9EG, UK

T: +44 (0) 1590 613 500  
F: +44 (0) 1590 613 555  
E: anem@gill.co.uk

[www.gill.co.uk](http://www.gill.co.uk)



Copyright © Gill Instruments 2011

Gill Instruments Ltd  
Reg No. 3154453 Registered Office: The George Business Centre, Christchurch Road, New Milton, BH25 6QJ


# GILL

## Gill Instruments Limited

### Certificate of Conformance

This is to certify that:

Product: **1390-75-B-313** **WO75 ENH HTD CON PAN MT LG**  



Serial Number: **1351005**  


Has been manufactured within Gill Instruments Limited Quality Management System, approved to the requirements of BS EN ISO 9001.

The instrument has been tested and calibrated with equipment having full traceability to national standards where applicable and meets the requirements of the Operating Manual Specification and Electrical Conformity (EC Declaration in Manual).

The instrument comes with a **12 Month** warranty against defective materials or workmanship from the date of purchase, providing it has not been tampered with and has been returned through an authorised route to Gill Instruments Limited.

Date created: 20 Dec 2013

Signed by:   
.....  
Les Rann  
Quality Engineer



Gill Instruments Limited  
Saltmarsh Park  
67 Gosport Street  
Lymington  
Hampshire  
SO41 9EG, UK

Tel: +44 (0)1590 613500  
Fax: +44 (0)1590 613501  
E-mail: [gill@gill.co.uk](mailto:gill@gill.co.uk)  
Web: [www.gill.co.uk](http://www.gill.co.uk)



View our current Anemometer range here