



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA  
Website: <http://www.seabird.com>

Phone: (425) 643-9866  
FAX: (425) 643-9954  
Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 1/8/2014 Description SBE-3 Temperature Sensor

### SBE Sensor Information:

Model Number: 03

Serial Number: 6042

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test:

Tested By: VG

