



# Sea-Bird Electronics, Inc.

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866  
Fax: (425) 643-9954  
Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: 11/13/14

Description: SBE-9plus CTD

### Sensor Information:

Model Number: 09

Serial Number: 1213

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 15 Minutes

Passed Test: Yes

Tested By: ap

High pressure is generally equal to the maximum depth rating of the instrument

