



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St, Bellevue Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 Email: seabird@seabird.com

Digiquartz Pressure Calibration dP/dT Corrected Coefficients

(Changed coefficients are posted in italics)

Pressure Transducer Serial Number: 132931

Original Calibration Date: 2014-11-03

Date of Correction: 2014-11-18

Installed in: SBE 9Plus S/N 1213

PRESSURE COEFFICIENTS

<i>C1</i>	<i>-41326.81</i>	<i>psia</i>
<i>C2</i>	<i>-4.4584627e-01</i>	<i>psia/deg C</i>
<i>C3</i>	<i>1.3315e-02</i>	<i>psia/deg C²</i>

D1	0.034889
D2	0.0

<i>T1</i>	<i>30.3213</i>	<i>μsec</i>
<i>T2</i>	<i>-4.098794e-04</i>	<i>μsec/deg C</i>
T3	3.9678e-06	μsec/deg C ²
T4	3.64375e-09	μsec/deg C ³
T5	0e+00	

AD590M =
AD590B =
Slope = 1.0
Offset = 0.0

Corrected at Sea-Bird Electronics as per Paroscientific Calibration and Sea-Bird Electronics dP/dT tests. The original calibration from Paroscientific assumes an operating temperature range of 0 to 125 degrees C. dP/dT correction adjusts this operating range to a nominal range of 0 to 22 degrees C. This increases the accuracy of the transducer in this temperature range.

NOTE: Original coefficients from Paroscientific are attached to this form for informational purposes and should not be used.

CALIBRATION COEFFICIENTS

SERIAL NO : 132931

PRESSURE TRANSDUCER

DATE : 11-04-2014

MODEL : 410K-134	PRESSURE RANGE : 0 to 10000 psia	TEMP. RANGE : 0 to 125 deg C	PORT :
---------------------	-------------------------------------	---------------------------------	--------

PRESSURE COEFFICIENTS

U = temperature
(deg C)

$$C = C_1 + C_2U + C_3U^2$$

$$D = D_1 + D_2U$$

$$T_0 = T_1 + T_2U + T_3U^2 + T_4U^3 + T_5U^4$$

T = pressure period
(μsec)

Pressure : (psia)

$$P = C \left(1 - \frac{T_0^2}{T^2}\right) \left(1 - D \left(1 - \frac{T_0^2}{T^2}\right)\right)$$

C ₁	-41324.33 psia
C ₂	-4.66874E-01 psia/deg C
C ₃	1.33150E-02 psia/deg C ²

D ₁	0.034889
D ₂	0

T ₁	30.32211 μsec
T ₂	-4.17047E-04 μsec/deg C
T ₃	3.96781E-06 μsec/deg C ²
T ₄	3.64375E-09 μsec/deg C ³
T ₅	0

(11-03-2014)

PAROSCIENTIFIC, INC.

4500 148th AVENUE N.E.
REDMOND, WA. 98052

CUSTOMER : SEABIRD ELECTRONICS, INC.

SALES ORDER : 31634

PREPARED BY : WMR



CALIBRATION COEFFICIENTSSERIAL NO : **132931**

PRESSURE TRANSDUCER

DATE : **11-04-2014**

MODEL : 410K-134	PRESSURE RANGE : 0 to 10000 psia	TEMP. RANGE : 0 to 125 deg C	PORT :
----------------------------	--	--	--------

PRESSURE COEFFICIENTS AT FIXED TEMPERATURE

(only valid at specified temperature)

T = pressure period (μsec)Pressure equation : (psia)

$$P = C \left(1 - \frac{T_0^2}{T^2} \right) \left(1 - D \left(1 - \frac{T_0^2}{T^2} \right) \right)$$

Temperature: **21.0 C**

C (psia)	-41328.26				
D	0.034889				
T ₀ (μsec)	30.31514				

(11-03-2014)

PAROSCIENTIFIC, INC.4500 148th AVENUE N.E.
REDMOND, WA. 98052CUSTOMER : **SEABIRD ELECTRONICS, INC.**SALES ORDER : **31634**PREPARED BY : **WMR**