

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0455
CALIBRATION DATE: 23-Dec-15

SBE 45 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 2.401494e-005
a1 = 2.758593e-004
a2 = -2.542089e-006
a3 = 1.546935e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	686253.7	1.0000	0.0000
4.5000	585878.8	4.5000	-0.0000
15.0000	371787.0	15.0000	-0.0000
18.5000	321486.3	18.5000	0.0000
24.0000	257363.6	24.0001	0.0001
29.0001	211526.4	28.9999	-0.0002
32.5000	185002.4	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

