



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0399  
 CALIBRATION DATE: 09-Dec-17

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.791122e-001                      CPcor = -9.5700e-008  
 h = 1.431878e-001                      CTcor = 3.2500e-006  
 i = -4.973873e-004                      WBOTC = 6.9909e-007  
 j = 5.762149e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2623.26	0.00000	0.00000
1.0000	34.7464	2.97056	5270.86	2.97056	0.00000
4.5000	34.7270	3.27713	5471.15	3.27713	0.00001
15.0000	34.6851	4.25722	6066.39	4.25720	-0.00001
18.4999	34.6764	4.60180	6261.93	4.60180	-0.00000
24.0000	34.6670	5.15887	6565.37	5.15887	0.00001
29.0000	34.6620	5.67988	6836.60	5.67990	0.00002
32.5000	34.6593	6.05170	7023.50	6.05169	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

