



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0385
 CALIBRATION DATE: 28-Dec-17

SBE 45 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 3.975706e-005
 a1 = 2.734631e-004
 a2 = -2.339709e-006
 a3 = 1.498227e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	667071.0	1.0000	0.0000
4.5000	569562.0	4.5000	-0.0000
15.0000	361545.9	15.0000	0.0000
18.4999	312664.8	18.4999	-0.0000
24.0000	250339.6	24.0000	0.0000
29.0000	205780.7	29.0000	-0.0000
32.5000	179994.4	32.5000	0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature

