



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0385
 CALIBRATION DATE: 28-Dec-17

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.831415e-001 CPcor = -9.5700e-008
 h = 1.332936e-001 CTcor = 3.2500e-006
 i = -1.032830e-004 WBOTC = 2.7605e-007
 j = 2.642849e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2716.70	0.00000	0.00000
1.0000	34.7606	2.97166	5442.50	2.97166	0.00000
4.5000	34.7411	3.27833	5648.77	3.27833	-0.00000
15.0000	34.6994	4.25879	6262.03	4.25879	0.00000
18.4999	34.6906	4.60348	6463.52	4.60348	-0.00000
24.0000	34.6806	5.16067	6776.29	5.16067	0.00000
29.0000	34.6744	5.68169	7055.88	5.68169	-0.00000
32.5000	34.6702	6.05339	7248.52	6.05330	-0.00009

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

