

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
www.sea-birdscientific.com

## C-Star Calibration

Date **4.12.18** S/N# **CST-1856DR** Pathlength **25 cm**

---

	Analog output	Digital output
$V_{\text{dark}}$	<b>0.002 V</b>	<b>0 counts</b>
$V_{\text{air}}$	<b>4.814 V</b>	<b>15806 counts</b>
$V_{\text{ref}}$	<b>4.700 V</b>	<b>15432 counts</b>

Temperature of calibration water **22.9 °C**  
Ambient temperature during calibration **22.1 °C**

---

Relationship of transmittance ( $T_r$ ) to beam attenuation coefficient ( $c$ ), and pathlength ( $x$ , in meters):  $T_r = e^{-cx}$

To determine beam transmittance:  $T_r = (V_{\text{sig}} - V_{\text{dark}}) / (V_{\text{ref}} - V_{\text{dark}})$

To determine beam attenuation coefficient:  $c = -1/x * \ln(T_r)$

$V_{\text{dark}}$  Meter output with the beam blocked. This is the offset.

$V_{\text{air}}$  Meter output in air with a clear beam path.

$V_{\text{ref}}$  Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain  $V_{\text{ref}}$ .

Ambient temperature: meter temperature in air during the calibration.

$V_{\text{sig}}$  Measured signal output of meter.