



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 1410  
CALIBRATION DATE: 20-Jun-18

### SBE 18 pH CALIBRATION DATA

#### pH COEFFICIENTS:

pHslope = 4.6177  
pHoffset = 2.5147

pH BUFFER	TEMPERATURE (°C)	INSTRUMENT OUTPUT (volts)	COMPUTED pH	RESIDUAL (pH)
4.0	23.9	1.697	3.996	-0.004
7.0	23.9	2.517	7.009	0.009
10.0	23.9	3.330	9.996	-0.004

Vout = Instrument pH sensor output in volts

$$\text{pH} = 7.0 + (\text{Vout} - \text{pHoffset}) / (\text{pHslope} * \text{°K} * 1.98416\text{E-}4)$$

Residual (pH) = instrument pH - buffer pH

