

**Wintronics, Inc.**  
50 Division Ave, Suite 1  
Millington, NJ 07946  
Phone: (908) 647-0144 Fax: (908) 647-8379

---

## Certificate of Calibration

Certificate No.: J0133081

---

Manufacturer: <b>Heitronics</b>	Description: <b>Infrared Thermometer</b>
Model No: <b>CT15.10</b>	Serial No: <b>10975</b>

---

Customer: U Of Alaska, Fairbanks Seward Marine Center 201 Railway Ave Seward, AK 99664	Customer PO: Customer Asset No:
Temperature (C): 23 Humidity (%): 40 Procedure: W60985, Rev 2	Technician: PLW Date Cal: 13-Jul-2018 Date Due: 13-Jul-2019

The manufacturer's specifications of the above instrument have been confirmed by comparison to standards which are regularly calibrated using accepted values of natural physical constants, ratio type self-calibrating techniques, comparison to standards which are traceable to NIST or other NMI, or compared to consensus standards. Wintronics' calibration procedures comply with MIL-STD-45662A & ISO-9001-2000. Wintronics' Quality System is accredited to ISO/IEC 17025:2005.

As received condition: Inoperative  
As shipped condition: Out of Tolerance  
Type of Calibration: Normal

### Calibration Standards

Manufacturer	Model	Description	Asset #	Calibration Date	Date Due	Cert. No.
Heitronics	TRT3.82	Transfer Infrared Pyrometer	W185	16-May-2018	31-May-2019	J0132126

  
**Peter Winter**  
President

Certified By

