



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0454  
 CALIBRATION DATE: 02-Mar-19

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -8.172293e+000      CPcor = -9.5700e-008  
 h = 2.605009e+000      CTcor = 3.2500e-006  
 i = -7.013399e-001      WBOTC = 2.4343e-006  
 j = 5.369998e-002

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2745.04	0.00000	0.00000
1.0000	34.8469	2.97833	5434.82	3.03721	0.05888
4.5000	34.8221	3.28522	5639.29	3.20296	-0.08226
15.0000	34.7796	4.26759	6248.11	4.29479	0.02720
18.5000	34.7700	4.61289	6448.18	4.94345	0.33056
23.9999	34.7596	5.17111	6431.88	4.88385	-0.28726
28.9999	34.7507	5.69277	6618.13	5.64663	-0.04614
32.5001	34.7422	6.06454	6704.42	6.06409	-0.00045

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

