

pHfit V 2.2
pH Calibration TimeStamp: 2019-10-08T01:21:06
SBE18/27 pH serial number: 18-1410

t = 292.150000
a = 0.000000
b = 7.605600
c = 4.875600
d = 18.000000
deter = 54.000000
slope = 0.270867
offset = 2.535200
final slope = 4.672755

pHslope = 4.6728
pHoffset = 2.5352

residuals:

pH	Vout	res
4.00	1.722	0.004
7.00	2.537	-0.007
10.00	3.347	0.004