



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2020-02-26**

Description: **SBE-18 pH Sensor**

### Sensor Information:

Replaced the main piston "O"-Rings.  
Replaced the pH Electrode.

Model Number: **SBE-18**

Serial Number: **1401**

### Pressure Test Protocol:

Low Pressure Test: **50** PSI Held For: **15** Minutes

High Pressure Test: **1500** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **BLT**

High pressure is generally equal to the maximum depth rating of the instrument

