



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0385
 CALIBRATION DATE: 21-Feb-20

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.878794e-001 CPcor = -9.5700e-008
 h = 1.340637e-001 CTcor = 3.2500e-006
 i = -1.590318e-004 WBOTC = 2.7605e-007
 j = 2.958155e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2716.70	0.00000	0.00000
1.0000	34.8475	2.97838	5439.00	2.97839	0.00001
4.5000	34.8272	3.28565	5645.04	3.28565	-0.00000
15.0000	34.7834	4.26800	6257.63	4.26797	-0.00003
18.5000	34.7736	4.61332	6458.92	4.61333	0.00001
24.0000	34.7626	5.17152	6771.34	5.17153	0.00001
29.0000	34.7557	5.69351	7050.65	5.69353	0.00002
32.5000	34.7509	6.06587	7243.10	6.06586	-0.00002

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

