

Calibration Date: 06/24/20

Job No.: L20087

Model Number: QSP2300

Serial Number: 70777

Operator: TPC

Standard Lamp: V-040(1/3/2019)

Operating Voltage Range: 6 to 15 VDC (+)

Note: The QSP2300 output is a voltage that is proportional to the log of the incident irradiance. To calculate irradiance, use this formula:

Irradiance = Calibration factor \* (10<sup>v</sup>Light Signal Voltage - 10<sup>v</sup>Dark Voltage)

Dry Calibration Factor: 4.05E+12 quanta/cm<sup>2</sup>-sec per volt 6.72E-06 μEinsteins/cm<sup>2</sup>-sec per volt

Wet Calibration Factor: 7.15E+12 quanta/cm<sup>2</sup>-sec per volt 1.19E-05 μEinsteins/cm<sup>2</sup>-sec per volt

Sensor Test Data and Results<sup>2)</sup>

Sensor Supply Current (Dark): 3.4 mA

Supply Voltage: 6 Volts

Lamp Integrated PAR Irradiance: 9.40E+15 quanta/cm<sup>2</sup>-sec 0.01561 μEinsteins/cm<sup>2</sup>-sec

Immersion Coefficient: 0.566

Test Irrad.

Nominal Filter OD	Expected Transmission	Calibrated Trans.	Sensor Voltage	Expected Voltage	Voltage % Error	Measured Trans.	Transmission Error (%)	Test Irrad. (quanta/cm <sup>2</sup> -sec)
No Filter	100%	100.00%	3.366	3.366	0%	100.00%	0.0	9.40E+15
0.3	50%	36.10%	2.921	2.924	0%	35.82%	0.8	3.37E+15
0.5	32%	27.60%	2.806	2.807	0%	27.51%	0.3	2.59E+15
1	10%	9.27%	2.334	2.333	0%	9.24%	0.3	8.69E+14
2	1%	1.11%	1.421	1.411	1%	1.09%	1.7	1.03E+14
3	0.10%	0.05%	0.318	0.094	71%	0.05%	18.3	4.37E+12
RG780	0.00%	0.00%	0.247	0.013	95%	0.03%	-100.0	3.10E+12

Dark Before: 0.013 Volts  
 Light - No Filter Hldr.: 3.366 Volts  
 Dark After - NFH: 0.013 Volts  
 Average Dark 0.0134 Volts

Notes:

1. Annual calibration is recommended.

2) This section is for internal use and for more advanced analysis.