



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0454  
 CALIBRATION DATE: 24-Feb-22

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.008664e+000      CPcor = -9.5700e-008  
 h = 1.348079e-001      CTcor = 3.2500e-006  
 i = -2.302569e-004      WBOTC = 2.4343e-006  
 j = 3.514878e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2739.03	0.00000	0.00000
1.0000	34.6570	2.96364	5432.60	2.96363	-0.00001
4.4999	34.6376	3.26951	5637.21	3.26953	0.00001
15.0000	34.5958	4.24742	6245.72	4.24743	0.00001
18.5000	34.5867	4.59119	6445.68	4.59116	-0.00003
24.0000	34.5756	5.14676	6756.08	5.14678	0.00001
29.0000	34.5682	5.66624	7033.55	5.66624	-0.00000
32.5000	34.5615	6.03656	7224.61	6.03648	-0.00009

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

