

SUNA V2

Saltwater Calibration Verification

SUNA Serial Number:	1999
RMA Number:	

Reference Solution	RS-232 output		Analog output (V)		Analog output (mA)	
	Instrument Output	Reference Range	Instrument Output	Reference Range	Instrument Output	Reference Range
Pure Water (0 μm)	-0.312	0 μm (± 2)	0.394	0.28 \pm 0.1 (0.18–0.38)	4.752	4.76 \pm 0.35 (4.41–5.11)
Seawater (0 μm)	-0.096	0 μm (± 2)	0.429	0.28 \pm 0.1 (0.18–0.38)	4.725	4.76 \pm 0.35 (4.41–5.11)
Low Nitrate (40 μm)	38.890	40 μm (± 4)	1.915	1.81 \pm 0.16 (1.65–1.97)	10.563	10.86 \pm 0.61 (10.25–11.47)

Calibration File Name	Date	Technician Initials
SNA1999A	11/1/2022	PF

Final Reference Update Needed? (Yes / No):	No
--	-----------

Calibration File Name	Date	Technician Initials