



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0385
 CALIBRATION DATE: 01-Feb-23

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.891600e-001 CPcor = -9.5700e-008
 h = 1.350358e-001 CTcor = 3.2500e-006
 i = -5.321246e-004 WBOTC = 2.7605e-007
 j = 5.804611e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2716.75	0.00000	0.00000
1.0000	34.6258	2.96123	5432.53	2.96130	0.00007
4.5000	34.6055	3.26679	5638.18	3.26674	-0.00004
15.0000	34.5616	4.24366	6249.64	4.24354	-0.00012
18.5000	34.5518	4.58705	6450.54	4.58703	-0.00002
24.0000	34.5402	5.14208	6762.29	5.14230	0.00022
29.0000	34.5300	5.66068	7040.37	5.66057	-0.00011
32.5000	34.5183	6.02987	7230.89	6.02822	-0.00165

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

