



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2023-02-23**

Description: **SBE-45 Mini Thermosalinograph**

Sensor Information:

Replaced the main piston "O"-Rings.

*The unit is not designed for high pressure applications.

Model Number: **SBE-45**

Serial Number: **399**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **30** Minutes

High Pressure Test: **0** PSI Held For: **0** Minutes

Passed Test: **True**

Tested By: **JK**

High pressure is generally equal to the maximum depth rating of the instrument

