



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 4791
 CALIBRATION DATE: 13-Jul-23

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.74304045e+000
 h = 1.25668215e+000
 i = -7.03507529e-004
 j = 1.10639892e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.78564	0.00000	0.00000
-1.0001	34.5878	2.78787	5.47267	2.78788	0.00001
0.9999	34.5882	2.95831	5.59513	2.95831	-0.00000
14.9999	34.5884	4.24659	6.44551	4.24658	-0.00002
18.4999	34.5882	4.59135	6.65456	4.59136	0.00001
29.0000	34.5850	5.66868	7.26876	5.66870	0.00001
32.5000	34.5734	6.03841	7.46778	6.03840	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / 10 (1 + δ * t + ε * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

