



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2023-10-19**

Description: **SUNA-V2**

### Sensor Information:

Model Number: **SUNA-V2**

Serial Number: **2182**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **750**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **PP**

**High pressure is generally equal to the maximum depth rating of the instrument**

