



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2023-11-16**

Description: **SUNA-V2**

### Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SUNA-V2**

Serial Number: **1999**

### Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **150** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **aP**

