

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration and Repairs

Date 8/19/2024 Customer
S/N# CST-1856DR Work order 1005521493

Standard Service

- Cleaned meter
- Replaced all seals
- Focused and aligned optics
- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test
- Performed water calibration
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet
- Check for a zinc anode

Additional Repairs

BAA-540771 X1
EXA-KX0169 X1
MAA-110018 X1
GXA-HX0007 X1
MAA-802035 X1

Comments