

In Coriolis 9525  
May '26 DDN



**TELEDYNE RESON**  
Everywhereyoulook™

## SVP Test and Calibration certificate

SVP Type :	SVP70
SVP Serial No.	2824016

Date of issue : 21/10/2024

Temperature Calibration :	Hart 1504 s/n B38892 & Thermistor s/n 3702
Point 1:	4.5 °C
Point 2:	16.5 °C
Point 3:	25.5 °C
Pressure Calibration :	Custom Built Tank (TestUnit ASF150 Ser# 41-10-0007-R03)
Point 1:	0 Bar
Point 2:	277.4 Bar
Point 3:	544.6 Bar

### RMS Speed of Sound Errors

Temperature Validation :	0.0126 m/s
Pressure Validation :	0.1232 m/s

Calibration & Final Function Test :

Sign :

23 OKT. 2024

QA Signature :

Inits : PBH OK



**TELEDYNE RESON**  
Everywhereyoulook™

TELEDYNE-RESON A/S, Fabriksvangen 13, DK-3550 Slangerup  
Fax: +45 4738 0066, Phone: +45 4738 0022