



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866
Fax: (425) 643-9954
Email: seabird@seabird.com

Pressure Test Certificate

Test Date: 06/11/14

Description: SBE-43 DO Sensor

Sensor Information:

Model Number: 43

Serial Number: 2945

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 15 Minutes

Passed Test: Yes

Tested By: RH

High pressure is generally equal to the maximum depth rating of the instrument

