

SBE49 FastCAT CTD Sensor

Instrument Configuration

Instrument Serial Number: Instrument Firmware Version: Zero Conductivity Frequency: Communications Format: Communications Settings: 49-0636 7.2.5 2675.60 RS232 9600 baud, 8 Data Bits, No Parity

Installed Devices/Sensors

Data Format	Measurement	Sensor Type	Serial Number	Rating
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure Sensor	Druck	12351123	350m(350 dBar)

Maximum Depth:

350m

CAUTION - The maximum deployment depth will be limited by the measurement range of the pressure sensor, if installed, an attached sensor, if installed, or the housing.

www.seabird.com | seabird@seabird.com | Tel: +1 425 643 9866

Copyright 2017 Sea-Bird Scientific. All rights reserved.