



**SEA-BIRD**  
SCIENTIFIC

## SBE49 FastCAT CTD Sensor

### Instrument Configuration

Instrument Serial Number: 49-0720  
Instrument Firmware Version: 1.3a  
Zero Conductivity Frequency: 2585.15  
Communications Format: RS232  
Communications Settings: 9600 baud, 8 Data Bits, No Parity

### Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A
Frequency	Conductivity	Internal	N/A	N/A
Count	Pressure Sensor	Druck	12421551	350m(350 dBar)

Maximum Depth: **350m**

**CAUTION** - The maximum deployment depth will be limited by the measurement range of the pressure sensor, if installed, an attached sensor, if installed, or the housing.



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0720  
 CALIBRATION DATE: 16-Jun-24

SBE 49 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 7.867946e-004  
 a1 = 2.971306e-004  
 a2 = -3.780368e-006  
 a3 = 2.517069e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
0.9999	753831.051	0.9999	0.0000
4.5000	668097.576	4.4999	-0.0001
15.0000	449759.034	15.0001	0.0001
18.5000	389746.983	18.5000	-0.0000
24.0000	307299.644	24.0000	-0.0000
29.0000	243776.051	29.0000	-0.0000
32.5000	205080.102	32.5000	0.0000


n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV \* 2.295e+010 + 9.216e+008) / (6.144e+004 - MV \* 5.3e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln<sup>2</sup>(R)] + a3[ln<sup>3</sup>(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0720  
 CALIBRATION DATE: 16-Jun-24

SBE 49 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.810187e-001                      CPcor = -9.5700e-008  
 h = 1.474725e-001                      CTcor = 3.2500e-006  
 i = -4.054746e-004  
 j = 5.514128e-005


BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2585.15	0.0000	0.00000
0.9999	34.7015	2.96708	5185.06	2.9671	0.00001
4.5000	34.6813	3.27324	5381.73	3.2732	-0.00001
15.0000	34.6373	4.25197	5966.24	4.2520	-0.00001
18.5000	34.6276	4.59603	6158.24	4.5960	0.00002
24.0000	34.6168	5.15222	6456.17	5.1522	0.00001
29.0000	34.6100	5.67232	6722.42	5.6723	-0.00003
32.5000	34.6033	6.04303	6905.73	6.0430	0.00001

f = Instrument Output (Hz) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) =  $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0720  
 CALIBRATION DATE: 14-Jun-24

SBE 49 PRESSURE CALIBRATION DATA  
 508 psia S/N 12421551

COEFFICIENTS:

PA0 =	1.025158e-001	PTCA0 =	5.234009e+005
PA1 =	1.543162e-003	PTCA1 =	5.179836e+000
PA2 =	4.165187e-012	PTCA2 =	-9.527906e-002
PTEMPA0 =	-5.920771e+001	PTCB0 =	2.499388e+001
PTEMPA1 =	5.389609e+001	PTCB1 =	1.754386e-004
PTEMPA2 =	-3.378881e-001	PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.58	532862.0	1.5	14.59	0.00	32.50	1.72	533562.31
113.10	596665.4	1.6	113.06	-0.01	29.00	1.65	533568.78
211.64	660508.8	1.6	211.62	-0.01	24.00	1.56	533567.90
310.28	724403.8	1.6	310.29	0.00	18.50	1.46	533561.85
408.96	788281.4	1.6	408.97	0.00	15.00	1.39	533550.25
507.70	852151.0	1.6	507.67	-0.01	4.50	1.19	533515.95
409.30	788510.6	1.6	409.33	0.00	1.00	1.13	533504.15
310.58	724606.0	1.6	310.60	0.00			
211.82	660640.4	1.6	211.82	-0.00			
113.18	596741.2	1.6	113.17	-0.00	TEMPERATURE (°C)		SPAN
14.58	532859.8	1.6	14.59	0.00	-5.00	24.99	
					34.90	25.00	

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$


$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 14-Jun-24 0.00





Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2024-06-13**

Description: **SBE-49 FastCat Sensor**

### Sensor Information:

Model Number: **SBE-49**

Serial Number: **0720**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **500**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **s.a**

**High pressure is generally equal to the maximum depth rating of the instrument**

