



Paroscientific, Inc.
 4500 148th Ave. N.E.
 Redmond, WA 98052, USA
 Tel: (425) 883-8700
 Fax: (425) 867-5407
 www.paroscientific.com

PACKING SLIP # 40594-01
 11/24/25
 Page 1 of 1
 SALES ORDER # 40594

SHIP TO: UNIVERSITY OF ALASKA FAIRBANKS SEWARD MARINE CENTER ATTN: GABE MATTHIAS 201 RAILWAY AVE SEWARD , AK 99664	BILL TO: 6699 UNIVERSITY OF ALASKA A/P PO BOX 730 SEWARD, AK 99664
PO NUMBER.: PH DT: 10/16/2025	FOB.....: REDMOND, WA USA
ORDER DATE: 10/16/25	SHIPPED VIA : FEDEX
	SHIP CHARGES:

LINE\PART NBR.....\PICK..\WH LOC.....\SHIP..\REFN\#BOXES\WEIGHT
 ITEM\DESCRIPTION.....\QTY.....\LOT...\QTY\QTY

1 9777 1
 SERVICE DEPT. -- NON WARRANTY REPAIR

INST: MODEL MET4A, P/N 1564-301, S/N 146581.
 REPAIR AND CALIBRATE PER SR# 13186.

CUSTOMER INSURES

ECCN: EAR99. NLR. THESE COMMODITIES, TECHNOLOGY, OR SOFTWARE
 MUST BE EXPORTED IN ACCORDANCE WITH THE EXPORT ADMINISTRATION
 REGULATIONS. DIVERSION CONTRARY TO U.S. LAW PROHIBITED.
 COUNTRY OF ORIGIN: USA

Paroscientific
 QA/Shipping
 Inspection

SHIP FEDEX 2DAY# 222392004 11

ENCLOSURES AS CHECKED: AJ
 1) BILL OF LADING
 2) SCD
 3) COEFFICIENTS AJ

PACKAGED BY: [Signature]
 CARTON _____ OF _____

All sales are subject to Paroscientific's terms and conditions, a copy of which may be found at <http://www.paroscientific.com/pdf/TandC1.pdf>

Paroscientific, Inc.

4500 148th Ave. NE
Redmond, WA 98052
Tel: (425) 883-8700 Fax: (425) 867-5407

Service Report Number: 13186

Rev. Initials: CDD **Rev. Date** 11/19/2025

Date Received: 10/9/2025

Tel: 401-207-6632

Company: UNIVERSITY OF ALASKA

Fax/Email: grmatthias@alaska.edu

Serial Number	Part Number	Model Number	Warranty	NCR	Original Zero Date	Customer Comment/ Reason for Return
146581	1564-301	MET4A	<input type="checkbox"/>		8/8/2019	for calibration.

Customer Supplied Box <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO	External Condition <input type="checkbox"/> LIKE NEW <input checked="" type="checkbox"/> FAIR <input type="checkbox"/> POOR Comments: <input style="width: 100%;" type="text"/>	Offset Error (ppm) <input style="width: 50px;" type="text" value="26"/> Current Draw (mA) <input style="width: 50px;" type="text" value="173"/> Firmware Version <input style="width: 50px;" type="text" value="M3.10"/>
Accessories Returned <input type="checkbox"/> Cable <input type="checkbox"/> Manual Other: <input style="width: 50px;" type="text" value="none"/>	Pressure Port <input type="checkbox"/> BLOCKED <input type="checkbox"/> PARTIALLY BLOCKED <input checked="" type="checkbox"/> CLEAR <input type="checkbox"/> N/A-ACCELEROMETER Comments: <input style="width: 100%;" type="text"/>	Oil Level <input type="checkbox"/> FULL <input type="checkbox"/> OIL VISIBLE <input type="checkbox"/> LOW OIL <input type="checkbox"/> NO OIL VISIBLE <input checked="" type="checkbox"/> N/A
MET Products RH <input style="width: 50px;" type="text" value="54.37"/> TEMP <input style="width: 50px;" type="text" value="18.99"/>		Elec. Isolation: <input type="checkbox"/> PASS <input type="checkbox"/> FAIL <input checked="" type="checkbox"/> N/A

Analysis: The exterior surfaces of the unit have debris, paint is chipping and signs of corrosion. There is black tape that covers the LED panel. An incoming functional test was performed. Both RS-232/485 communications were established. The relative error of the pressure sensor was 26ppm when compared to a reference laboratory barometer. The aspirator fan is functional. The unique status setting is MD=8.

Recommendations: Wipe clean unit. Perform zero adjust calibration to the barometer which includes NIST certificate. Return with unique status setting MD=8. \$1,500

Revisions: During rezero, it was found the intelligent board needs replaced. Wipe clean unit. Replace intelligent board. Perform zero adjust calibration to the barometer which includes NIST certificate. Return with unique status setting MD=8.

Date Analysis Completed: 10/9/2025 **Completed By:** M. Hao

*Quotation is valid for one month from the "Date Analysis Completed." Products remaining at Paroscientific after 2 months will be returned and an evaluation fee will be charged for each product.

CUSTOMER COPY

Paroscientific, Inc.
Pressure Instrument Configuration

SN: 146581 Part Number: 1564-301 Model: MET4A Port:
 Calibration Date: 21-Nov-25 Report No: 28466 Technician: WMR
 Pressure Range: 500 to 1100 hPa Temperature Range: -40 to +60 deg C

Customer: University of Alaska Report Date: 21-Nov-25
 Address : 201 Railway Ave. P730 Sales Order: 40594
 Seward, AK 99664 USA S/R Number : 13186

Configuration		Calibration Coefficients	
BL: 0	PT: N	U0: 5.806784 μsec	
BR: 9600	QD: -	Y1: -4041.64 deg C / μsec	
DD: -	QO: -	Y2: -9129.143 deg C / μsec^2	
DL: 0	SL: 0	Y3: 0 deg C / μsec^3	
DM: 0	SN: 146581	C1: 96.48165 psi	
DO: 0	ST: 10	C2: 4.277946 psi / μsec	
DP: 6	SU: 0	C3: -106.6632 psi / μsec^2	
ID: 01	TI: 666	D1: 0.0340303	
IM: -	TR: 952	D2: 0.0000000	
LL: -	TU: 0	T1: 27.79397 μsec	
LH: -	UF: 1.000000	T2: 0.612522 $\mu\text{sec} / \mu\text{sec}$	
MC: Y	UL:	T3: 20.86147 $\mu\text{sec} / \mu\text{sec}^2$	
MD: 8	UM: user	T4: 0 $\mu\text{sec} / \mu\text{sec}^3$	
MN: MET4A	UN: 3	T5: 0.000 $\mu\text{sec} / \mu\text{sec}^4$	
OP: -	US: 0	TC: 0.6781676	
PF: 1.103161	VR: M3.10	PA: 0.0000000	
PI: 666	ZI: 0	PM: 1.0000000	
PL: 19.00000	ZS: 0		
PO: 0	ZL: 0		
PR: 238	ZV: .000000		
PS: 0			

Met4/Met4A Coefficients		
GD: 0	JC: .0000000	LN: 0
GT: 0	JD: .0	LO: 00:00:00
PC: -	JE: 0	LT: .0000000
HY: 2	JF: .0	LW: 0
H1: 20268722	JG: 0	PX: 3
S1: 1.000000	JJ: 0	
S2: 1.000000	JK: 0	
Z1: 0	JS: .0	
Z2: 0	JU: 0	

Paroscientific, Inc.
 4500 148th Ave. N.E. Redmond, WA 98052
 Phone: (425)883-8700 Fax: (425)867-5407
 Web: <http://www.paroscientific.com>
 Email: support@paroscientific.com

Prepared by





Paroscientific, Inc.
Precision Pressure Instrumentation

Certificate of Conformance

Customer: UNIVERSITY OF ALASKA FAIRBANKS
Purchase Order: PH DT: 10/16/2025
Digiquartz® Model: MET4A
Part Number: 1564-301
Serial Number(s): 146581

Paroscientific, Inc. certifies that the part(s) identified above complies with the requirements of the above order and has been manufactured in accordance with engineering drawings, material and process specifications, testing procedures, and applicable specification drawing of Paroscientific, Inc.

The Digiquartz® model(s) identified has been calibrated and tested over the specified pressure and temperature range and meets the requirements of the applicable specification control drawing. Primary pressure, temperature, and transfer standards used at Paroscientific, Inc. for calibration and testing have traceability to the National Institute of Standards and Technology (NIST) and the SI through a National Metrology Institute, radiometric techniques, or natural physical constraints and are regularly checked and calibrated according to Paroscientific quality procedures and in accordance with the requirements of ISO 9001:2015.

Paroscientific
QA/Shipping
Inspection

11

**Quality Assurance
Shipping Inspection Stamp**

Date: 11/24/2025



Primary pressure and temperature standards used in the calibration and testing of Paroscientific Digiquartz® models have traceability to the National Institute of Standards and Technology and to the SI through the documentation below.

Calibration Instruments:

Bell and Howell Primary Pressure Standard:

Bell and Howell, Model 6-201-0001, Piston/Cylinder P2-919/C2-1523 via DH Calibration Report No. 15441, 16653 and 1500190591. Weight Set 1, P/N 6-002-0002, via DH Calibration Report No. 14481, 16654, 1284473284 and 1500190311. Weight Set 2, P/N 6-002-0002, via DH Calibration Report No. 31227, 39628, 68390, 1317389777, 1500156263, 1500195122 and 1500238176-1. Piston/Cylinder P2-d652/C2-1378 via DH Instruments Calibration Report No. 31226, 39627, 68389, 1317739617, 1500156259, 1500195119 and 1500238175. Piston/Cylinder P1-231/C1-384 via DH Instruments Calibration Report No. 13170 and 1284475131. Piston/Cylinder P/N 6-201, No. P1-949/C1-922, via DH Instruments Calibration Report 17176 and 17445.

DH Primary Pressure Standard, Oil Operated Gauge:

DH Instruments, Model 5306, Piston/Cylinder S/N 3375, via DH Calibration Certificate Report No. 32354, 45306, 66563, 1305698573, 1500163455 and 1500217726. Piston/Cylinder 3511 via DH Calibration Report No. 32353, 45307, 66562, 1305625084, 1500163363 and 1500217811. Mass Set S/N 2032 via DH Calibration Report No. 45305, 45308, 1305538925, 1305530461, 150016339, 1500163394 and 150021768.

DH Primary Pressure Standard, Pneumatic Operated Gauge:

DH Instruments, Model 5203, Piston/Cylinder S/N 4845, via DH Calibration Certificate No. 38275, 1300177141, 1500162809 and 1500217494. Mass Set S/N 2032/3293 via DH Calibration Certificate No. 45305, 1300200369, 1500162832 and 1500217766.

DH Primary Pressure Standard, Pneumatic Operated Gauge:

DH instruments, Model PG7601 via DH Instruments Calibration No. 69127, 1500172365 and 1500278487. Piston/Cylinder S/N 305 via DH Instruments Calibration No. 69125, 1500172367, 1500278543. DH Instruments 35 kg Mass Set No. 2052 and Bell No. 261A via DH Instruments Calibration Report No. 69126, 69124, 1500172362, 1500172363, 1500277884 and 1500277905.

Hygroclip S3 MET4/4A Part number 1560-XXX and 1561-XXX:

Humidity and Temperature calibrations are traceable to NIST through Rotronic Instrument Corporation; 160 E. Main Street, Huntington, NY 11743.

Hygroclip HC2-S3 MET4/4A Part number 1563-XXX and 1564-XXX Swiss Calibration Service (SCS)

Humidity and Temperature calibrations are traceable to SCS through Rotronic AG Grindelstrasse 6 8303 Bassersdorf Phone: 044-838-1111 E-mail: info@rotronic.ch



Digiquartz® Model: MET4A

Serial Number(s): 146581

The Paroscientific Digiquartz® model identified above has been calibrated and tested with one or more of the following primary pressure standards. All have traceability to the National Institute of Standards and Technology (NIST) and to the SI.

Bell and Howell Primary Pressure Standard

Pneumatic Absolute or Gauge Dead Weight Tester Part Number: 6-201-0001, S/N 4034 and SN 1014

- | | |
|---|--|
| <input type="checkbox"/> Piston/Cylinder: 6-001-0002, P2-919/C2-1523
Weight Set 1: 6-002-0002
Range: 1.5 to 50 psi [10 to 345 kPa]
Accuracy: 0.010 percent of reading | <input type="checkbox"/> Piston/Cylinder: 6-001-0001, P2-949/C1-922
Weight Set 2: 6-002-0002
Range: 0.3 to 5 psi [2 to 34 kPa]
Accuracy: 0.015 percent of reading |
| <input checked="" type="checkbox"/> Piston/Cylinder: 6-001-0002, P2-652/C2-1378
Weight Set 1: 6-002-0002
Range: 1.5 to 50 psi [10 to 345 kPa]
Accuracy: 0.010 percent of reading | |

DH Primary Pressure Standard

Pneumatic Absolute or Gauge Dead Weight Tester Part Number: PG7601 S/N 161

- Piston/Cylinder: S/N 305, Mass Set: S/N 2052
Range: 0.7 to 50 psi [5 to 345 kPa] absolute mode, 0.29 to 50 psi [2 to 345 kPa] gauge mode
Accuracy: 0.002 percent of reading

DH Primary Pressure Standard

Pneumatic Gauge Dead Weight Tester, Model 5203, S/N 5557

- Piston/Cylinder: S/N 4845, Mass Sets: S/N 2032, S/N 3293
Range: 20 to 1600 psi [0.14 to 11 MPa]
Accuracy: 0.005 percent of reading

DH Primary Pressure Standard

Pneumatic Gauge Dead Weight Tester, Model 5306, S/N 3505

- Piston/Cylinder: S/N 3375, Mass Sets: S/N 2032
Range: 40 to 20000 psi [0.3 to 138 MPa]
Accuracy: 0.01 percent of reading above 200 psi [1.4 MPa] or 0.02 psi [0.14 kPa] at lower pressure
- Piston/Cylinder: S/N 3511, Mass Sets: S/N 2032
Range: 145 to 72500 psi [1 to 500 kPa]
Accuracy: 0.02 percent of reading above 725 psi [5 MPa] or 0.145 psi [1 kPa] at lower pressure

Test Stamp



Date: 11/24/2025