



**SEA-BIRD**  
SCIENTIFIC

Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2022-09-01**

Description: **Power Data Interface Module**

### Sensor Information:

Model Number: **PDIM**

Serial Number: **0289**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **1000**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **SG**

High pressure is generally equal to the maximum depth rating of the instrument

