



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2024-02-20**

Description: **SBE-38 Digital Thermometer**

### Sensor Information:

Replaced the main piston "O"-Rings.  
Replaced 4-pin connector.

Model Number: **SBE-38**

Serial Number: **0668**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **10000**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **AP**

High pressure is generally equal to the maximum depth rating of the instrument

