



Sea-Bird Electronics, Inc.
 13431 NE 20th Street
 Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518024

03-APR-2023

320496974

SERVICE REPORT

Service Request

Date

Sales Order

CUSTOMER INFORMATION

Name: UNIVERSITY OF ALASKA - SEWARD
 Account : 40280223
 ETHAN ROTH
 EHROTH@ALASKA.EDU
 858-232-1848

PO Number:

Bill To Address

UNIVERSITY OF ALASKA - SEWARD
 SEWARD MARINE CENTER / SFOS
 P.O. BOX 730
 SEWARD,AK,99664,US

Ship To Address

UNIVERSITY OF ALASKA - SEWARD
 SEWARD MARINE CENTER / SFOS
 201 RAILWAY AVENUE
 SEWARD,AK,99664,US

PRODUCT INFORMATION

Item: 18.121

Item Description: SBE18 pH SENSOR, 1200M STRAIGHT ENDCAP, MCBH

Serial: 18-1401

Special Notes

Services Requested:

SERVICE18

CUSTOMER NOTES: not working properly

Problems Found:

The pH electrode has reached the end of its life expectancy

Problems Found:

The pH electrode was found to have reached the end of its life expectancy.

Services Performed:

Perform initial diagnostic evaluation.

Replaced the pH electrode.

Replaced the O-rings.

Performed a hydrostatic pressure test.

Performed a "Final" Calibration.

Item	Item Description	Qty
SERVICE18	SERVICE OF SBE 18 pH SENSOR INCLUDES CALIBRATION, REPLACEMENT OF pH ELECTRODE, O-RINGS AND HYDROSTATIC PRESSURE TEST IF DEEMED NECESSARY. ALSO INCLUDES TYPICAL REPAIRS. (Does not include: Major electronic board replacement. Instruments deemed "Damaged Beyond Repair". Missing/damaged mounts or cables. Upgrades, Modifications or Spares.	1

Unbilled Items

Item	Item Description	Qty
------	------------------	-----



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518024

03-APR-2023

320496974

SERVICE REPORT

Service Request

Date

Sales Order

24023P	pH electrode (prepped for installation, document 67072)	1
--------	---	---