



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1410
CALIBRATION DATE: 09-Sep-25

SBE 18 pH CALIBRATION DATA

pH COEFFICIENTS:

pHslope = 4.6252

pHoffset = 2.6143

pH BUFFER	TEMPERATURE (°C)	INSTRUMENT OUTPUT (volts)	COMPUTED pH	RESIDUAL (pH)
4.0	23.6	1.794	3.988	-0.012
7.0	23.6	2.621	7.024	0.024
10.0	23.6	3.428	9.988	-0.012

Vout = Instrument pH sensor output in volts

$pH = 7.0 + (V_{out} - pH_{offset}) / (pH_{slope} * ^\circ K * 1.98416E-4)$

Residual (pH) = instrument pH - buffer pH

