



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0399  
 CALIBRATION DATE: 08-Feb-24

SBE 45 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 8.003005e-005  
 a1 = 2.686654e-004  
 a2 = -2.032297e-006  
 a3 = 1.411797e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	643644.6	1.0000	0.0000
4.5000	549119.1	4.5000	-0.0000
15.0000	347769.4	15.0001	0.0001
18.5000	300531.3	18.4999	-0.0001
24.0000	240356.3	24.0000	0.0000
29.0000	197379.5	29.0000	-0.0000
32.5000	172530.1	32.5000	0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature

