



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 4790  
 CALIBRATION DATE: 23-Jul-24

SBE 4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.77797893e+000      CPcor = -9.5700e-008 (nominal)  
 h = 1.31638388e+000      CTcor = 3.2500e-006 (nominal)  
 i = 1.62133701e-004  
 j = 5.49121208e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.72454	0.00000	0.00000
-1.0001	33.8690	2.73525	5.30610	2.73522	-0.00003
1.0000	33.8685	2.90254	5.42419	2.90257	0.00004
15.0000	33.8681	4.16739	6.24451	4.16739	-0.00001
18.4999	33.8660	4.50567	6.44612	4.50565	-0.00002
28.9999	33.8521	5.56185	7.03831	5.56189	0.00004
32.5000	33.8246	5.92227	7.22917	5.92225	-0.00002

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h \* f<sup>2</sup> + i \* f<sup>3</sup> + j \* f<sup>4</sup>) / 10 (1 + δ \* t + ε \* p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

