

Seatex MGC Calibration Certificate



KONSGBERG

| | |
|---------------------------------|----------------|
| Seatex MGC model number: | R3.v |
| Serial number: | 50998 |
| Calibration certificate number: | 20231125-50998 |

MGC


1. Roll and Pitch Accuracy Tests

| Roll and pitch accuracy | Test requirement | Roll | Pitch |
|----------------------------------|------------------|--------|--------|
| RMS static roll and pitch [deg] | 0.01 | Passed | Passed |
| RMS dynamic roll and pitch [deg] | 0.01 | Passed | Passed |

The static accuracy was measured by sampling at 4 Hz for 30 minutes, when the Seatex MGC is stationary.

The dynamic accuracy was measured in a rate table test with simultaneous sinusoidal excitation in two axes for 10 minutes.

Plots of results from dynamic test of Seatex MGC with serial number 50998



2. Rate Gyro Accuracy Tests

| Angular rate accuracy | Test requirement | R-axis | P-axis | Y-axis |
|--|------------------|--------|--------|--------|
| RMS rate sensor noise [deg/s] | 0.01 | Passed | Passed | Passed |
| RMS rate sensor scale factor error [%] | 0.001 | Passed | Passed | Passed |

The angular rate sensor noise level was measured by sampling at 4 Hz for 30 minutes, when the Seatex MGC is stationary.

The rate gyro scale factor error was tested by single-axis rotations on a rate table at $\pm 30^\circ/\text{s}$ and at $\pm 50^\circ/\text{s}$.

3. Accelerometer Accuracy Tests

| Linear acceleration accuracy | Test requirement | R-axis | P-axis | Y-axis |
|--|------------------|--------|--------|--------|
| STD acceleration sensor noise [m/s^2] | 0.0002 | Passed | Passed | Passed |
| RMS acceleration sensor scale factor error [%] | 0.008 | Passed | Passed | Passed |

The acceleration sensor noise level was measured by sampling at 4 Hz for 30 minutes, when the Seatex MGC is stationary.

The accelerometer scale factor was measured by tilting the Seatex MGC in steps of 90° around a circle.

The Calibration Certificate test requirements are the technical specification limits in the MGC User Manual

A two-axis rate table with temperature chamber (DC 246-TCM from Acutronic AG, 1998) was used to test the unit.

All tests were performed at room temperature according to test procedures in the MGC Production Manual