13431 NE 20th ST,
Bellevue, 98005, WA

(425) 643-9866

www.sea-birdscientific.com

C-Star Calibration

Date July 16, 2026 S/N# CST-1681DR Pathlength 25cm

	Analog output	Digital output
V_d	0.003 V	0 counts
V_{air}	4.729 V	15536 counts
V_{ref}	4.702 V	15447 counts

Temperature of calibration water	20.56	°C
Ambient temperature during calibration	20.5	°C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters): $\mathbf{Tr = e^{-cx}}$

To determine beam transmittance: $\mathbf{Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})}$

To determine beam attenuation coefficient: $\mathbf{c = -1/x * ln\ (Tr)}$

$\mathbf{V_d}$ Meter output with the beam blocked. This is the offset.
$\mathbf{V_{air}}$ Meter output in air with a clear beam path.
$\mathbf{V_{ref}}$ Meter output with clean water in the path.
Temperature of calibration water: temperature of clean water used to obtain V_{ref}.
Ambient temperature: meter temperature in air during the calibration.
$\mathbf{V_{sig}}$ Measured signal output of meter.